

SCIENTIFIC REPORTS



OPEN

Multi-dimensional TOF-SIMS analysis for effective profiling of disease-related ions from the tissue surface

Received: 25 November 2014

Accepted: 13 May 2015

Published: 05 June 2015

Ji-Won Park^{1,2,7,*}, Hyobin Jeong^{3,*}, Byeongsoo Kang⁴, Su Jin Kim^{1,5}, Sang Yoon Park⁶, Sokbom Kang⁶, Hark Kyun Kim⁶, Joon Sig Choi⁵, Daehee Hwang^{3,4,7} & Tae Geol Lee^{1,2}

Time-of-flight secondary ion mass spectrometry (TOF-SIMS) emerges as a promising tool to identify the ions (small molecules) indicative of disease states from the surface of patient tissues. In TOF-SIMS analysis, an enhanced ionization of surface molecules is critical to increase the number of detected ions. Several methods have been developed to enhance ionization capability. However, how these methods improve identification of disease-related ions has not been systematically explored. Here, we present a multi-dimensional SIMS (MD-SIMS) that combines conventional TOF-SIMS and metal-assisted SIMS (MetA-SIMS). Using this approach, we analyzed cancer and adjacent normal tissues first by TOF-SIMS and subsequently by MetA-SIMS. In total, TOF- and MetA-SIMS detected 632 and 959 ions, respectively. Among them, 426 were commonly detected by both methods, while 206 and 533 were detected uniquely by TOF- and MetA-SIMS, respectively. Of the 426 commonly detected ions, 250 increased in their intensities by MetA-SIMS, whereas 176 decreased. The integrated analysis of the ions detected by the two methods resulted in an increased number of discriminatory ions leading to an enhanced separation between cancer and normal tissues. Therefore, the results show that MD-SIMS can be a useful approach to provide a comprehensive list of discriminatory ions indicative of disease states.

Time-of-flight secondary ion mass spectrometry (TOF-SIMS) has been used to study surface properties of specimens due to its high molecular specificity and surface sensitivity^{1–3}. TOF-SIMS spectra represent chemical compositions of positive and negative secondary ions (small molecules) in the very near surface region. Recently, TOF-SIMS has been used to study the surface of complex biological samples such as cells^{4–6} and tissues^{7–10}. Furthermore, TOF-SIMS can be used to characterize secondary ions from the surface of tissue samples. Comparative analysis of the detected ions between disease and normal samples can reveal the ions indicative of disease states^{4,9}. For example, Yun *et al.*⁹ catalogued secondary ions from seven gastric cancer and eight normal tissues using TOF-SIMS and found 66 ions discriminating gastric cancer tissues from normal ones. Such studies show that TOF-SIMS is emerging as a promising tool to identify the ions that reflect disease states from the surface of patient tissues.

¹Center for Nano-Bio Measurement, Korea Research Institute of Standards and Science, Daejeon, Republic of Korea. ²Department of Nano and Bio Surface Science, University of Science and Technology, Daejeon, Republic of Korea. ³School of interdisciplinary bioscience and bioengineering, POSTECH, Pohang, Republic of Korea. ⁴Department of New Biology, DGIST, Daegu, Republic of Korea. ⁵Department of Biochemistry, Chungnam National University, Daejeon, Republic of Korea. ⁶National Cancer Center, Goyang, Republic of Korea. ⁷Center for Systems Biology of Plant Senescence and Life History, Institute for Basic Science, Daegu, Republic of Korea. *These authors contributed equally to this work. Correspondence and requests for materials should be addressed to D.H. (email: dhwang@dgist.ac.kr) or T.G.L. (email: tglee@kriss.re.kr)

In TOF-SIMS analysis, enhancing the ionization of small molecules from the surface of a specimen is important to increase the number of detected ions, and several methods have been developed to improve ionization capability. First, Wu *et al.*¹¹ proposed a matrix-enhanced SIMS analysis where a matrix (2,5-dihydroxybenzoic acid) used commonly in matrix assisted laser desorption ionization (MALDI) is deposited on the surface of the specimen, followed by TOF-SIMS analysis. In this analysis, ion-beam sputtering of the solid mixtures of analytes and matrix resulted in an increased production of high mass ions. Second, Linton *et al.*¹² introduced metal-assisted SIMS (Meta-SIMS) analysis where Ag is deposited on the low-density poly(ethylene) surface, followed by TOF-SIMS analysis. They found that Ag deposition resulted in an increased emission of secondary ions from the surface. Recently, several studies^{13–15} have further demonstrated that submonolayer depositions of silver or gold on the surface improve the ionization of surface molecules. However, how these methods increase the number of detected ions and thereby improve the efficiency for identifying disease-related ions has not been systematically investigated.

Here, we present a multi-dimensional TOF-SIMS (MD-SIMS) approach for effective profiling of the secondary ions from the surface of tissue samples, which can lead to an efficient identification of disease-related ions at the tissue surface. This approach first uses conventional TOF-SIMS to analyze the fresh surface of the tissues, after which Au is deposited on the surface, followed by TOF-SIMS analysis (i.e., Meta-SIMS). This serial analysis provides multi-dimensional data generated by both the conventional TOF-SIMS and Meta-SIMS. To demonstrate the utility of a MD-SIMS analysis, we analyzed ovarian cancer and adjacent normal tissues sequentially using TOF- and Meta-SIMS. The comparison of the TOF- and Meta-SIMS data revealed shared and distinct ions detected by the two SIMS methods, indicating that particular signals were suppressed and enhanced by Meta-SIMS. By combining the ions detected by TOF- and Meta-SIMS, the MD-SIMS analysis produced an increased number of discriminatory ions between cancer and normal tissues, which can be used as indicators of disease states.

Results

MD-SIMS analysis of ovarian cancer and normal tissues. For a comparative analysis of the secondary ions at the tissue surface, we collected the tissues from 10 ovarian cancer patients. The characteristics of the patients are summarized in Supplementary Table S1. For each of the 10 ovarian tissues, we first performed two serial sectioning. One section was used for H&E staining to identify cancer and adjacent normal regions, whereas the other section was used for TOF- and Meta-SIMS analyses (Fig. 1a). Based on the H&E staining, we selected 40 different cancer and adjacent normal regions (e.g. boxes in Fig. 1a) that contain high densities of cancer and adjacent normal cells, respectively, from the 10 ovarian cancer tissues (**Methods**). Multiple areas in each tissue were analyzed to account for the heterogeneity arisen from the varying amounts of cancer cells (see **Discussion**). For each of the 80 selected regions, we performed TOF-SIMS analysis (Fig. 1b, left box). Subsequently, we performed Meta-SIMS analysis after depositing Au on the same regions for which TOF-SIMS analysis was performed (Fig. 1b, right box). The Meta-SIMS analysis was performed for 50 of the 80 selected regions (30 cancer and 20 adjacent normal regions). In each SIMS analysis, we analyzed the secondary ions with both positive and negative modes. Thus, these analyses resulted in 80 positive and negative TOF-SIMS spectra and 50 positive and negative Meta-SIMS spectra (Fig. 1c).

Comparison of TOF-SIMS with Meta-SIMS data. For each dataset (80 TOF-SIMS spectra or 50 Meta-SIMS spectra), we identified the ion peaks and then aligned the peaks using TOFSIMS-P as previously described⁹. Of these ion peaks, we focused on the peaks that were detected in more than 50% of ovarian cancer and normal tissue samples in each dataset (i.e., 20 cancer and 20 normal samples in TOF-SIMS datasets; 15 cancer and 10 normal samples in Meta-SIMS dataset), as previously reported⁹. The 50% cutoff was used to ensure statistical reliability in the following analyses of the peaks including the comparison of TOF- and Meta-SIMS and the identification of discriminatory ions between ovarian cancer and normal tissues (see **Discussion**). Based on this criterion, we identified 301 positive and 331 negative ion peaks from the 80 TOF-SIMS spectra, and 388 positive and 571 negative ion peaks from the 50 Meta-SIMS spectra (Fig. 2a).

Of these peaks, 426 were detected commonly by both TOF- and Meta-SIMS analyses (Fig. 2b, 1st pie chart). Of the 426 shared peaks, 250 were increased in their intensities by Meta-SIMS analysis, compared to the TOF-SIMS analysis, indicating enhanced ionization of the corresponding ions by Au deposition (Fig. 2c, top). In contrast, 176 showed decreased intensities in Meta-SIMS analysis (Fig. 2c, bottom). This suggests that Au deposition can also suppress the ionization of these ions. On the other hand, 739 were detected uniquely by TOF-SIMS (206 peaks) or Meta-SIMS analysis (533 peaks), respectively (Fig. 2b, 1st pie chart). The 533 peaks uniquely detected by Meta-SIMS analysis comprise 1) Au adducts (189 peaks) for which the same ions were detected at m/z -196.97 (Au) by TOF-SIMS analysis (Fig. 2d, top) and 2) the 344 peaks detected truly uniquely by Meta-SIMS analysis (Fig. 2d, bottom) (Fig. 2b, 2nd pie chart). These data indicate that the two SIMS analyses provide complementary information for the compositions of the secondary ions from the tissue surface by detecting shared and distinct ions with different intensities.

Discriminatory ions between ovarian cancer and normal tissues. To identify the ions indicative of disease states, we next identified the peaks that discriminated the ovarian cancer samples from the




Figure 1. Overview of the MD-SIMS framework.

normal samples. To this end, we first combined positive and negative ion peaks detected from TOF- and MetA-SIMS analyses. From all of these detected peaks, we used only the 632 and 959 peaks in the TOF- and MetA-SIMS datasets, respectively, (Fig. 2a), which were detected in more than 50% of ovarian cancer and normal tissue samples to ensure their statistical power as discriminatory peaks (see **Discussion**). We then applied two-tailed T-test and also partial least square-discriminant analysis (PLS-DA) to these peaks. Of them, we finally selected 99 discriminatory peaks from TOF-SIMS data and 47 discriminatory peaks from MetA-SIMS data with $P < 0.01$ using two-tailed T-test and $VIP > 1$ obtained from PLS-DA, as previously described⁹ (Fig. 3a; Supplementary Table S2; **Methods**). Among these discriminatory peaks, only one was identified commonly from the TOF- and MetA-SIMS data, while 98 and 46 were identified uniquely in TOF- and MetA-SIMS data, respectively (Fig. 3b). The uniquely identified discriminatory




Figure 2. Shared and distinct ions detected by TOF-SIMS and Meta-SIMS analyses. (a). Relationships among positive (top) or negative (bottom) peaks measured by TOF-SIMS and Meta-SIMS analyses. (b). Proportion of the peaks commonly (shared) and uniquely (unique) detected by TOF-SIMS and Meta-SIMS analyses. The second pie chart shows the composition of the peaks uniquely detected by Meta-SIMS analysis. (c). Examples of the peaks detected commonly by TOF- and Meta-SIMS analyses with enhanced and suppressed ionizations in Meta-SIMS analysis, compared to those in TOF-SIMS analysis. (d). Examples of an Au adduct peak and a peak uniquely detected by Meta-SIMS analysis.

peaks indicate that MD-SIMS analysis can provide a more comprehensive view of the discriminatory signatures between ovarian cancer and normal samples than a single analysis of TOF- or Meta-SIMS. Of the 145 discriminatory peaks, four example peaks showing up- or down-regulated in ovarian cancer tissues, compared to adjacent normal tissues, are shown in Fig. 3c,d.

Comprehensive signatures to enhance discrimination between ovarian cancer and normal tissues. Compared to the conventional TOF-SIMS analysis, the additional Meta-SIMS analysis in MD-SIMS analysis increased the number of discriminatory ions (Fig. 3b). The extended list of the discriminatory peaks can provide a better discrimination between ovarian cancer and adjacent normal tissues. To test this notion, we applied PLS-DA to the 99 discriminatory peaks (Fig. 3b) identified by the conventional TOF-SIMS analysis and also multi-block PLS-DA (MPLS-DA) to the 145 peaks (Fig. 3b) identified by MD-SIMS analysis. PLS-DA showed that 68.4% of the separation between ovarian cancer and normal samples was achieved by the 99 discriminatory peaks identified by TOF-SIMS analysis. With




Figure 3. Analysis of TOF- and MetA-SIMS datasets. (a). Analytical framework for identification of discriminatory ions. (b). Relationships between the discriminatory ions identified from TOF- and MetA-SIMS datasets. (c–d). Up- or down-regulation of four example discriminatory ions in ovarian cancer tissues detected by TOF- (c) and MetA-SIMS analyses (d), compared to normal tissues.

these 99 peaks, a partial overlap between the ovarian cancer and normal samples was observed in the PLS latent space (Fig. 4a). In contrast, MPLS-DA showed that the use of the additional discriminatory peaks identified by MD-SIMS analysis improved the separation to 82.6%, resulting in a reduced overlap between the ovarian cancer and normal samples, compared to that observed from PLS-DA using only the discriminatory peaks identified by TOF-SIMS analysis (Fig. 4b). This result indicates that MD-SIMS analysis provided a comprehensive list of molecular signatures that can better discriminate ovarian cancer samples from normal samples.

Discussion

Although TOF-SIMS is a promising tool to identify the secondary ions reflecting disease states at the surface of patient tissues, its use has been hampered by the limited capacity of detecting the ions. Thus, several methods have been developed to improve the capability to ionize the ions at the surface. This has led to an increased number of the detected secondary ions, but the usefulness of these methods to efficiently identify the ions indicating disease states has not been systematically explored. In this study, we proposed MD-SIMS analysis, a combination of TOF-SIMS and MetA-SIMS analyses. By applying MD-SIMS analysis to ovarian cancer and adjacent normal tissues, we demonstrated that MD-SIMS analysis resulted in a more comprehensive set of the ions associated with disease states, compared to the conventional TOF-SIMS analysis.

A cancer tissue often shows heterogeneous molecular signatures across different regions of the tissue. The variation in the relative amounts of cancer and normal cells across the different regions mainly accounts for the heterogeneity. Different amounts of normal cells in the regions can dilute disease-related signatures with different extents. In this study, given the heterogeneity, to identify the ions that represent disease states, we performed TOF-SIMS analysis in the following manner. First, for each of 10 cancer tissues, we performed TOF-SIMS analyses of 4 different regions ($100 \times 100 \mu\text{m}^2$ per region) that contain high densities of the cancer cells based on the H&E staining, assuming that the 4 regions can provide sufficient information of cancer-related signatures in each tissue. For comparative analysis, we further performed 10 different regions for each of the 4 cancer tissues containing high densities of adjacent normal cells. These analyses resulted in the 40 cancer and normal datasets. Compared with the numbers of the regions analyzed during the discovery phase in other SIMS and metabolomics studies^{9,16,17}, the 40 cancer regions can be considered to be sufficiently large to provide meaningful cancer-related signatures. Second, we then identified the discriminatory ions as the ones that showed consistent differences




Figure 4. Integrated analysis of TOF- and Meta-SIMS datasets. 3-d PLS-DA score plots for TOF-SIMS (a) and MD-SIMS (b). LV1-3, 1st to 3rd PLS latent variables (LVs). The tables show that MD-SIMS enhanced the power of discrimination of ovarian cancer samples (red) from normal samples (blue), compared to TOF-SIMS. X-block, intensities of the 99 discriminatory peaks in TOF-SIMS data; X1- and X2-blocks, intensities of the 145 discriminatory peaks in TOF- and Meta-SIMS data, respectively; Y-block, binary vector in which ones and zeros represent ovarian cancer and normal samples, respectively. A percent variance captured by an LV in X- or Y-blocks indicates the amount of the information in X-block used to explain the amount of the separation explained by the LV among the binary values in Y-block.

in their abundances between the 40 cancer and normal datasets. This helps us focus on identification of the discriminatory ions that can represent the ovarian cancer states.

In this study, we performed MD-SIMS analysis under a static condition ($<10^{12}$ primary ions/cm²), assuming that no significant surface damage occurs, as previously reported¹⁸. Even under this condition, however, organic systems, such as cells or tissues, might be damaged during TOF-SIMS, and Meta-SIMS spectra can include the ions resulting from such damage, thus affecting the accuracy of the MD-SIMS analysis. To examine this issue, we performed two serial TOF-SIMS analyses under a static condition (Case 1) for a region of an ovarian tissue ($100 \times 100 \mu\text{m}^2$). In the MD-SIMS analysis, for a region of an ovarian tissue, we performed TOF-SIMS analysis followed by Meta-SIMS analysis (Case 2). In Case 1, high correlations (correlation coefficients close to 1) were observed between the intensities of the peaks detected by two serial TOF-SIMS analyses (Supplementary Fig. S1a), indicating that despite potential tissue damage, two serial TOF-SIMS analyses generated highly reproducible sets of the detected ions. In contrast, in Case 2, relatively smaller correlations (correlation coefficients = 0.771 and 0.717 for positive and negative modes, respectively) were observed between the intensities of the peaks detected by TOF- and Meta-SIMS analyses (Supplementary Fig. S1b), indicating that despite potential tissue damage, Meta-SIMS analysis could generate a reasonable level of the peaks independent of those detected by the initial TOF-SIMS analysis.

To identify discriminatory ions, in this study, we used the peaks detected in more than 50% of ovarian cancer and normal tissue samples. The peaks less frequently observed could be either chemical noises arisen from contamination or the peaks for real secondary ions that existed in a small number of the regions or tissues analyzed. The discriminatory peaks observed in a small number of samples can be also meaningful. In this study, however, among the detected peaks, we focused on the peaks that were detected in more than a certain percentage of ovarian cancer and normal tissue samples in each dataset, as previously reported⁹, which can ensure statistically reliable comparison of TOF- and Meta-SIMS and identification of statistically reliable discriminatory ions that are likely to be detected in new tissues by MD-SIMS. To determine the percentage cutoff, we varied the cutoff from 10% to 90% and found, as expected, that a high percentage cutoff decreases the number of the peaks analyzed, increasing the false negative rate in identification of discriminatory ions, whereas a small cutoff increases the number of the peaks, increasing the false positive rate (Supplementary Fig. S2). These data revealed no evidence that a particular percentage cutoff should be used. Thus, we chose 50% as a percentage cutoff such that the conclusions from the following analyses are likely to be valid at least in half of the samples to be analyzed.

Few approaches have been proposed to identify the molecules for the ions detected by TOF-SIMS^{19–21}, which often require significant amounts of efforts and time. In this study, our primary goal was to

propose MD-SIMS to effectively extend the list of the disease-related ions, rather than to demonstrate biological meanings of these ions by mapping them into molecules and pathways. Thus, we performed no experiments to identify the 145 discriminatory ions. However, it would be important to assign molecules to the discriminatory ions in order to assess whether the extended list of the discriminatory ions by MD-SIMS would be useful. We thus employed a previously reported approach that compares the masses of the 145 discriminatory ions with those of the TOF-SIMS ions for which the molecules were previously identified using the experimental approaches mentioned above²². As a result, we identified the molecules for 22 of the 145 discriminatory ions (Supplementary Table S3), including fatty acids (FAs), cholesterol (CH), glycerophospholipids (GP), and phosphatidylethanolamines (PE). CH, PE, and GP were identified from TOF-SIMS, and FAs were additionally identified from Meta-SIMS, showing that the integrated MD-SIMS analysis provided a more comprehensive set of molecules altered in ovarian cancer.

Based on the molecules assigned for the 22 discriminatory ions, we further built a pathway model describing altered lipid metabolism in ovarian cancer (Supplementary Fig. S3). In this model, GP and PE were increased in ovarian cancer tissues, compared to adjacent normal tissues, whereas FA and CH were decreased. Several of these changes have been previously reported to be associated with ovarian cancers. For example, altered expression of enzymes [FA synthase (FASN)²³, stearoyl-CoA desaturase-1 (SCD1)²⁴, and lysophosphatidic acid acyltransferase- β (LPAAT- β)²⁵] involved in the GP synthetic pathway was proposed as a potential indicator of the poor clinical outcome of ovarian cancer patients. Also, altered expression of HMG-CoAR(HMGCR), which regulates cholesterol synthesis, was proposed as a predictor of recurrence free survival in ovarian cancer²⁶. Moreover, in various cancers, GP or PE, components of plasma membrane, are increased during cell proliferation^{27,28}, consistent with up-regulation of GP or PE in ovarian cancer detected by MD-SIMS. All these data support potential validity of the extended discriminatory ions by MD-SIMS. However, detailed functional studies should be performed to demonstrate the validity of the discriminatory ions, together with validation of the discriminatory ions in large-scale cohorts. Furthermore, only a limited number of ovarian cancer samples ($n = 10$) was used in this study. Thus, the validity of the discriminatory ions should be tested in large-scale validation experiments, together with the detailed functional studies.

Principal component analysis (PCA) has been applied to the TOF-SIMS data to effectively explore the data structure in an unsupervised manner²⁹. In this study, however, we only applied PLS-DA to select the discriminatory ions based on the VIPs of the ions, together with the T-test. Thus, to examine whether PCA can provide additional knowledge of the data structure, we performed PCA using the 1,591 ion peaks (632 from TOF-SIMS and 959 from Meta-SIMS) to which PLS-DA was applied to select the discriminatory ions based on their VIPs. The first three PCs captured 54.5% (22.22, 18.97, and 13.30% respectively) of the total variance in the data (Supplementary Fig. S4a). PC1 (22.22%) showed no correlation with the difference between ovarian cancer and normal tissue samples, suggesting that the data contained a high level of variability independent of the difference between cancer and normal samples. However, PC2 (18.97%) began to show a significant ($P = 4.77 \times 10^{-4}$) correlation with such difference, indicating that the data also contained significant amount of information regarding the difference between cancer and normal samples (Supplementary Fig. S4b). To understand how PLS-DA handles these variabilities in the data independent of the difference between cancer and normal samples, we applied multi-block PLS-DA to the same data (Supplementary Fig. S4c). The comparison of both PCA and PLS-DA results showed that the application of PLS-DA to the 1,591 ion peaks allowed us to effectively focus on the information relevant to the difference between cancer and normal samples, thereby enabling correct identification of the discriminatory ions based on the VIPs. Moreover, multi-block PLS-DA further showed how the discriminatory ions identified from both TOF- and Meta-SIMS data additively contributed to capturing the difference between ovarian cancer and normal samples. In summary, MD-SIMS can be a useful approach to provide a more comprehensive list of the discriminatory ions that can be used to understand the disease states.

Methods

Sample preparation. Tissues were obtained from 10 ovarian cancer patients hospitalized at the National Cancer Center in Korea from 2005 to 2008 with informed consent in accordance with the guidelines approved by the institutional review board at the National Cancer Center (approval no. NCCNCS 09-309)³⁰. Samples were rapidly frozen in liquid nitrogen upon acquisition and then stored at -80°C before TOF-SIMS analysis. The samples were processed for TOF-SIMS analysis as previously described³¹. Briefly, two serial sections of each tissue of $10\ \mu\text{m}$ -thickness were generated at -20°C using a cryostat (Leica CM 3050S, Leica Microsystems Inc., IL). The 1st section was affixed to a glass slide and stained with hematoxylin and eosin (H&E) to identify the areas enriched with normal or cancer cells. The 2nd section was deposited onto a Si wafer that was sonicated with ethanol and acetone for 5 min each, rinsed with water, and then stored at -80°C until TOF-SIMS analysis. No chemical fixation was done because of the possibility that chemical artefacts could be detected by TOF-SIMS. Based on the H&E staining, we selected 4 different regions containing high densities of cancer cells from the 2nd section for each of the 10 ovarian cancer tissues (40 cancer regions). Also, we selected 40 adjacent normal regions from the ovarian tissues. To this end, we first selected 4 of 10 tissues containing high densities of adjacent normal cells, based on the H&E staining, and then further selected 10 adjacent normal regions from each of the 4 ovarian tissues (40 adjacent normal regions). Each selected region was analyzed by TOF-SIMS in both

the positive and negative modes. After the TOF-SIMS analysis, Au was deposited onto the tissue surface using a Quorum Technologies (Newhaven, East Sussex, U.K.) SC7640 sputter coater. The thickness of the gold was controlled to be 1 nm by a FT7607 quartz crystal microbalance stage and a FT7690 film thickness monitor. The samples were then analyzed by TOF-SIMS after metal deposition.

TOF-SIMS analysis. We used a TOF-SIMS V instrument (ION-TOF GmbH, Germany) equipped with a bismuth liquid metal ion gun (LMIG). A Bi_3^+ primary beam at 25 keV in a high-current bunched mode (pulse width = 0.65 ns and beam diameter = $3\ \mu\text{m}$) with a target current of 0.2 pA and a repetition rate of 5 kHz was used to obtain positive and negative spectra. The analysis area of $100 \times 100\ \mu\text{m}^2$ (128×128 pixels) was randomly rastered by the primary ions and were charge-compensated for the tissue samples by low-energy electron flooding. The primary ion dose density was maintained below 10^{12} ions/cm² to ensure a static SIMS condition. Mass resolution was higher than 7000 at $m/z < 500$ in both the positive and negative modes. Positive and negative ion spectra from each sample at specific areas were obtained. Positive and negative ion spectra were internally mass calibrated using CH_3^+ , C_2H_3^+ , C_3H_5^+ and $\text{C}_5\text{H}_{14}\text{NO}^+$ peaks, and CH^- , C_2H^- , C_4H^- and $\text{C}_{18}\text{H}_{35}\text{O}_2^-$ peaks, respectively. After the calibration, the resulting mass accuracy was 60 ppm for $m/z < 200$ and 150 ppm for $m/z > 200$ on average. Each spectrum was exported as an ASCII file.

Analysis of TOF-SIMS or Meta-SIMS data. From each of the spectra generated by TOF-SIMS, the peaks were automatically identified using the TOFBAT program in the built-in IonSpec software from ION-TOF. We searched for those peaks whose intensities were larger than 20 between 1 and 800 m/z . For each peak, the m/z , m/z range, and area of the peak were identified. Using the peaks from all the samples, we then applied a computational platform, TOFSIMS-P⁹, as previously reported for comparative analysis of the TOF-SIMS data obtained from cancer and normal samples, which involved 1) refinement and alignment of the peaks identified from all the samples and 2) selection of discriminatory ions. The same procedure was used to analyze the Meta-SIMS dataset.

Partial least square-discriminant analysis (PLS-DA). PLS-DA was applied to the combined set of positive and negative ion peaks detected from TOF-SIMS or Meta-SIMS analysis, as previously described⁹. Briefly, we defined X-block with the intensities of k peaks in n samples (i.e. $n \times k$ alignment matrix) and $n \times 1$ Y-block with information of sample classes (one for cancer and zero for normal). First, we estimated the missing values (less than 50% of intensities in the alignment table) using the k -nearest-neighbor imputing method³². Second, we selected the number of latent variables (LVs) as the one for which the mean error rate, computed by performing 10-fold cross-validations (CVs) 1000 times, reached the minimum as previously described⁹. Finally, we calculated the mean of variable importance on the projection (VIP) values for each peak by performing the 10-fold CV 1000 times using the determined number of PLS LVs. We selected the discriminatory ion peaks as the ones with the mean VIP > 1 ^{33,34}.

References

1. Belu, A. M., Graham, D. J. & Castner, D. G. Time-of-flight secondary ion mass spectrometry: techniques and applications for the characterization of biomaterial surfaces. *Biomaterials* **24**, 3635–3653 (2003).
2. Wang, H., Castner, D. G., Ratner, B. D. & Jiang, S. Probing the orientation of surface-immobilized immunoglobulin G by time-of-flight secondary ion mass spectrometry. *Langmuir* **20**, 1877–87 (2004).
3. Benninghoven, A. Chemical Analysis of Inorganic and Organic Surfaces and Thin Films by Static Time-of-Flight Secondary Ion Mass Spectrometry (TOF-SIMS). *Angew. Chemie Int. Ed. English* **33**, 1023–1043 (1994).
4. Kulp, K. S. *et al.* Chemical and biological differentiation of three human breast cancer cell types using time-of-flight secondary ion mass spectrometry. *Anal. Chem.* **78**, 3651–3658 (2006).
5. Nygren, K. A. N., Hagenhoff, B., Malmberg, P. E. R., Nilsson, M. & Richter, K. Bioimaging TOF-SIMS: High Resolution 3D Imaging of Single Cells. *Microsc. Res. Tech.* **974**, 969–974 (2007).
6. Fletcher, J. *et al.* A new dynamic in mass spectral imaging of single biological cells. *Anal. Chem.* **80**, 9058–9064 (2008).
7. Touboul, D., Halgand, F. & Brunelle, A. Tissue molecular ion imaging by gold cluster ion bombardment. *Anal. Chem.* **76**, 1550–1559 (2004).
8. Sjövall, P., Johansson, B. & Lausmaa, J. Localization of lipids in freeze-dried mouse brain sections by imaging TOF-SIMS. *Appl. Surf. Sci.* **252**, 6966–6974 (2006).
9. Yun, S. *et al.* TOFSIMS-P: A web-based platform for analysis of large-scale TOF-SIMS data. *Anal. Chem.* **83**, 9298–305 (2011).
10. McDonnell, L. *a et al.* Subcellular imaging mass spectrometry of brain tissue. *J. Mass Spectrom.* **40**, 160–8 (2005).
11. Wu, K. J. & Odom, R. W. Matrix-enhanced secondary ion mass spectrometry: a method for molecular analysis of solid surfaces. *Anal. Chem.* **68**, 873–82 (1996).
12. Linton, R. W. *et al.* Time-of-flight secondary ion mass spectrometric analysis of polymer surfaces and additives. *Surf. Interface Anal.* **20**, 991–999 (1993).
13. Delcorte, a. *et al.* Metal-Assisted Secondary Ion Mass Spectrometry Using Atomic (Ga + , In +) and Fullerene Projectiles. *Anal. Chem.* **79**, 3673–3689 (2007).
14. Altelaar, A., Klinkert, I. & Jalink, K. Gold-enhanced biomolecular surface imaging of cells and tissue by SIMS and MALDI mass spectrometry. *Anal. Chem.* **78**, 734–42 (2006).
15. Adriaensens, L., Vangaever, F. & Gijbels, R. Metal-assisted secondary ion mass spectrometry: influence of Ag and Au deposition on molecular ion yields. *Anal. Chem.* **76**, 6777–85 (2004).
16. Hansson, M. *et al.* Iodine content and distribution in extratumoral and tumor thyroid tissue analyzed with X-ray fluorescence and time-of-flight secondary ion mass spectrometry. *Thyroid* **18**, 1215–20 (2008).
17. Veselkov, K. *a et al.* Chemo-informatic strategy for imaging mass spectrometry-based hyperspectral profiling of lipid signatures in colorectal cancer. *Proc. Natl. Acad. Sci. U. S. A.* **111**, 1216–21 (2014).

18. Vickerman, J. C., Briggs, J. C. V. D. & Briggs, D. in *ToF-SIMS: Materials Analysis by Mass Spectrometry 2nd edn*, Ch. 1, 1–8 (IM Publications, 2013).
19. Sjövall, P., Lausmaa, J. & Johansson, B. Mass spectrometric imaging of lipids in brain tissue. *Anal. Chem.* **76**, 4271–4278 (2004).
20. Touboul, D., Brunelle, A. & Lapr votte, O. Structural analysis of secondary ions by post-source decay in time-of-flight secondary ion mass spectrometry. *Rapid Commun. Mass Spectrom.* **20**, 703–709 (2006).
21. Piehowski, P. D. *et al.* MS/MS methodology to improve subcellular mapping of cholesterol using TOF-SIMS. *Anal. Chem.* **80**, 8662–8667 (2008).
22. Hanrieder, J. & Ewing, A. G. Spatial Elucidation of Spinal Cord Lipid- and Metabolite- Regulations in Amyotrophic Lateral Sclerosis. *Sci. Rep.* **4**, 1–7 (2014).
23. Wang, H. Q. *et al.* Positive feedback regulation between AKT activation and fatty acid synthase expression in ovarian carcinoma cells. *Oncogene* **24**, 3574–3582 (2005).
24. Roongta, U. V. *et al.* Cancer Cell Dependence on Unsaturated Fatty Acids Implicates Stearoyl-CoA Desaturase as a Target for Cancer Therapy. *Mol. Cancer Res.* **9**, 1551–1561 (2011).
25. Diefenbach, C. S. M. *et al.* Lysophosphatidic acid acyltransferase-beta (LPAAT-beta) is highly expressed in advanced ovarian cancer and is associated with aggressive histology and poor survival. *Cancer* **107**, 1511–1519 (2006).
26. Brennan, D. J. *et al.* Tumour-specific HMG-CoAR is an independent predictor of recurrence free survival in epithelial ovarian cancer. *BMC Cancer* **10**, 125 (2010).
27. DeBerardinis, R. J., Lum, J. J., Hatzivassiliou, G. & Thompson, C. B. The biology of cancer: metabolic reprogramming fuels cell growth and proliferation. *Cell Metab.* **7**, 11–20 (2008).
28. Currie, E., Schulze, A., Zechner, R., Walther, T. C. & Farese, R. V. Cellular fatty acid metabolism and cancer. *Cell Metab.* **18**, 153–161 (2013).
29. Worley, B. & Powers, R. Multivariate Analysis in Metabolomics. *Curr. Metabolomics* **1**, 92–107 (2013).
30. Kang, S., Lee, A., Park, Y. & Lee, S. Alteration in lipid and protein profiles of ovarian cancer: similarity to breast cancer. *Int. J. Gynecol. Cancer* **21**, 1566–72 (2011).
31. Brunelle, A., Touboul, D. & Lapr votte, O. Biological tissue imaging with time-of-flight secondary ion mass spectrometry and cluster ion sources. *J. Mass Spectrom.* **40**, 985–99 (2005).
32. Albrecht, D., Kniemeyer, O., Brakhage, A. a & Guthke, R. Missing values in gel-based proteomics. *Proteomics* **10**, 1202–11 (2010).
33. Chan, C., Hwang, D., Stephanopoulos, G. N., Yarmush, M. L. & Stephanopoulos, G. Application of multivariate analysis to optimize function of cultured hepatocytes. *Biotechnol. Prog.* **19**, 580–98 (2003).
34. Chong, I.-G. & Jun, C.-H. Performance of some variable selection methods when multicollinearity is present. *Chemom. Intell. Lab. Syst.* **78**, 103–112 (2005).

Acknowledgements

This study was supported by the Converging Research Center for Mass Spectrometric Diagnosis (2013K000429), the Pioneer Research Center Program (NRF-2012-0009541), the Bio-Signal Analysis Technology Innovation Program (NRF-2006-2005074) and Global Frontier Project (H-GUARD_2013M3A6B2078962), Degree Research Center of KRCE, Proteogenomics program, Institute for Basic Science (CA1308), and POSCO Research Fund (Project NO. 2013Y008).

Author Contributions

D.H. and T.G.L. planned the projects and designed the experiments; J-W.P., S.J.K. and J.S.C. performed TOF-SIMS and Meta-SIMS experiment; J-W.P., H.J., B.K., D.H. and T.G.L. analyzed and interpreted the results; S.Y.P., S.K. and H.K.K. performed sample preparation; all authors participated in discussion and helped finalized the manuscript.

Additional Information

Supplementary information accompanies this paper at <http://www.nature.com/srep>

Competing financial interests: The authors declare no competing financial interests.

How to cite this article: Park, J.-W. *et al.* Multi-dimensional TOF-SIMS analysis for effective profiling of disease-related ions from the tissue surface. *Sci. Rep.* **5**, 11077; doi: 10.1038/srep11077 (2015).



This work is licensed under a Creative Commons Attribution 4.0 International License. The images or other third party material in this article are included in the article's Creative Commons license, unless indicated otherwise in the credit line; if the material is not included under the Creative Commons license, users will need to obtain permission from the license holder to reproduce the material. To view a copy of this license, visit <http://creativecommons.org/licenses/by/4.0/>